

Supplement

Title: 705 Inst.Sht. Supplement Issue: **4**
Print Date: August 1998 Issue Date: 5/25
Revision/Date: 2, 6/03 Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 705 & 707
CD Rev. & Date: 1, 2/2008
CD PN: 2088974

Change #1

Under **Specifications**, change the **Input/Output Protection**, to:

Input/output protection on mA range: self resetting, nonreplaceable 0.1 A fuse.

Change #2, 41512

Under **General Specifications**, following the **Weight** specification, add:

Protection Class: Pollution Degree II

Change #3

On the second panel, under **International Symbols**, add the following:

 N10140	Conforms to relative Australian standards.
---	--

Change #4, 57548

Under **General Specifications**, replace the **Size** with:

Size: 69.85 mm W x 142.87 mm L x 38.10 mm H
[2.75 in W x 5.625 in L x 1.50 in H]

With holster and Flex-Stand: 76.20 mm W x
158.75 mm L x 44.45 mm H [3.00 in W x 6.25 in
L x 1.75 in H]

Change #5, TPUBS-335

Under ***Specifications, DC Volts Input***,
replace the Accuracy with:

$\pm(0.0025 \% + 2)$