

Instrument Security Procedures

Model: 709, 709H, 710

Product Name: 709 Precision Loop Calibrator, 709H Precision Hart Loop Calibrator, 710 mA Loop Valve Tester

Instrument Description: 4-20 mA loop calibrator. The 709H/710 versions have HART communication capability.

Memory Description:

Microprocessor:

- 128 kB flash (709/709H)
- 256 kB flash (710)
- 4 kB EEPROM

Microprocessor RAM:

- 4 kB (709/709H)
- 8 kB (710)

Data logging/test data EEPROM:

- 256 K (709H/710 only)

Memory Cleaning Instructions:

Use Programming tool to erase microprocessor flash and EEPROM.

Use Command from keypad or PC to erase data logging memory (709H/710 only)