

Instrument Security Procedures

Model: Fluke 772/773

Product Name: Milliamp Process Clamp Meter

Instrument Description:

772 and 773 is an extended product category based on 771, targeting process maintenance, instrument maintenance, and engineers' troubleshooting instrumentation and control systems.

Memory Description:

The Fluke 772/773 use the same memory device: U44, micro controller, TI MSP430F449, 60kB Flash, 2kB RAM, and contains operating code for the product and calibration constants.

Memory Cleaning Instructions:

The user-controlled storage operation, other than calibration constants, is the storage of Hold. Hold is cleared when exiting the HOLD mode.